



ACTRIS NO_x QA workshop 2023

Anna Callaghan, James Lee, Lucy Carpenter and Katie Read

Wolfson Atmospheric Chemistry Laboratories University of York



Cape Verde location













Wind direction and wind speed







- NO_x measured via a custom, dual channel NO chemiluminescence instrument
 - Channel 1: NO, NO_x and background
 - NO₂ photolytically dissociated at 385 and 395 nm
 - Nafion dryer on sample
 - Channel 2: NO_v (currently not operational)








Two NO₂ converters used

 Blue light converter (BLC 2), high conversion efficiency (~ 92%) but thermal interferences due to BLC heating up

Diodes and quartz tube system (photolysis cell – PLC),

lower conversion efficiency (~ 54%), but minimal

interferences







NO_x Measurement cycle

- 5-minute measurement cycle
 - 2 minutes of NO measurement
 - 2 minutes of NO_x, 1 minute using BLC, 1 minute using PLC
 - 1 minute of background
- NO_x measured continuously at CVAO since 2006
 - Most recent BLC installed in 2015
 - PLC installed in 2017, diodes replaced in April 2022
- Limits of detection
 - NO \sim 0.8 ppt / hour
 - $-NO_2 \sim 2 ppt / hour$








NO_x Calibration

- Calibration every 61 hours
 - 3 NO sensitivity cycles
 - 2 NO₂ conversion efficiency cycles
- After calibration 30 minutes artifact measurement
 - PAG air used
 - For NO nighttime measurement used as artifact




NO Timeseries





NO Diurnal cycle - February 202





NO₂ Timeseries

